





ISO 9001/EN 29001/BS 5750 APPROVED BY BVQI

Lighting level is shown by digital display It is possible to measure illuminance from 0.1 to 19990 lux.



- 3 ranges changeable from low to high illuminance (200/2000/20000lux)
- Data hold function
- Digital light meter with separate light receiving sensor and meter.

Find Quality Products Online at:

www.GlobalTestSupply.com

sales@GlobalTestSupply.com

KYUHITSU ELECTRICAL INSTRUMENTS WORKS, LTD.

Digital light meter with separate light receiving sensor and meter.

■Specifications

Model	Model 5202				
Ranges	0.1~19990Lux				
Accuracy (23℃±5℃)	Lux	Accuracy	Fc	Accuracy	
	200	±(4% rdg+5 rdg)	20	±(4% rdg+5 rdg	
	2000	±(4% rdg+5 rdg)	200	±(4% rdg+5 rdg	
	20000	±(5% rdg+4 rdg)	2000	±(5% rdg+4 rdg)	
Current Consumption	2mA approx				
Response Time	2.5times / sec.				
Operating Temperature Range	0~50°C Below 80% RH				
Storage Temperature Range	-10°C~60°C				
Augular Incident light characteristics	30*Less than ±3%				
	60°Less than ±10%				
	80°Less than ±30%				
Spectral response characteristics	JIS C1609 CLASS A				
Power source	6F22(9V)				
Dimensions		Meter:148(L) × 71(W) × 36(H)mm			
	Light receiving sensor:85(L) \times 67(W) \times 32(H)mm				
Weight	270g				
Accessories	Ca	rrying Case × 1	Pho	otocall Cover × 1	
	(6F22(9V) × 1	Instri	uction Manual × 1	

■INSTRUMENT LAYOUT


1.Display window

2. Power Switch 4.Range Switch

3.Data hold switch

5.Off set adjustment 6.Light receiving sensor

7.Sensor Cover



DIGITAL ILLUMINOMETER



■Features

- Model 5201 is a highly portable and compact digital illuminance from 0.1 to 19,990 Lux, with auto range switching.
- The digital display is held for a preset time(about 20 seconds)and,therefore, facilitates reading, recording and measuring in any

■Specifications

MODEL			
Ranges	0.1~19990Lux(automatic 3 range switching)		
Accuracy	±5%rdg±3dgt		
Measuring Time	3 times per second		
Temperature humidity characteristics	±3%(at 20°C)		
Augular Incident Light Characteristics	Closely related to the Lambert's cosine law		
Spectral Response	Closely related to the spectral		
Characteristics	luminous efficiency (of a standard observer)		
Power Source -	6F22(9V)×1		
Dimensions	166(L)×68(W)×32(D)mm		
Weight	180g approx		
Accessories	Photocell Cover	6F22×1	
	Soft Carrying Case	Instruction Manual	



Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use.

Failure to follow the safety rules can cause fire, trouble, electrical shock, etc.

Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

%The contents of this catalogue may be subject to change without notice.

■DISTRIBUTOR



KYORITSU ELECTRICAL INSTRUMENTS WORKS,LTD.

No. 5-20 Nakane 2-chome, Meguro-Ku, Tokyo, 152-0031 Japan Phone:81-3-3723-0131 Fax:81-3-3723-0152 URL: http://www.kew-ltd.co.jp E-mail:info@kew-ltd.co.ip

www.GlobalTestSupply.com